





## Oxygen Concentrator

Model #:	Provider Name: Hart Medical Equipment
Serial #:	Phone #:
Date:	Contact Person:

# **Setting up your equipment:**

- Place unit in a well-ventilated area.
- Plug directly into a working grounded outlet.
- Attach cannula (Picture 1) to oxygen port on the unit.



#### Picture 1

## **Using your equipment:**

- Turn on the concentrator. It will alarm briefly to set internally.
- Ensure liter flow is set at prescribed level. **Do not** change liter flow setting without a prescription from your physician.
- Place cannula properly into nose and behind the ears. (Picture 2)
- You will notice a "poofing" noise every few seconds. This is the normal operation of your concentrator as it separates the oxygen from the room air gases that are not needed.



Picture 2

### Attaching a humidifier bottle (if needed):

- Remove supply-tubing adaptor.
- Secure humidifier to concentrator.
- Attach tubing to humidifier output. (Picture 3)
- Make sure the water in the humidifier bottle bubbles once unit is turned on. If bubbling does not occur, check the connections.



Picture 3

#### Maintenance:

- Clean the cabinet filters with warm soapy water at least once a week.
- Wipe the unit with a damp cloth using a mild disinfectant once a week or as needed.
- Internal maintenance should be performed only by your equipment provider.
- Oxygen purity should be checked regularly by your equipment provider.

#### Your local provider:





©Copyright The VGM Group B.L.U.E. Sheet images, drawings, and text courtesy of VGM contracted vendors.